



EU Declaration of Conformity
(in accordance with EN ISO/IEC 17050:2010)

Product:

CPEC300 (with CR6)
CPEC306 (with CR6)
CPEC310 (with CR6)

Manufacturer:

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F
www.campbellsci.com

This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):

Model	Description
CPEC300	Expandable Closed Path Eddy Covariance System w/EC155 & Pump Module
CPEC300-6NA	Expandable Closed Path Eddy Covariance System w/EC155 & Pump Module, with CR6
CPEC306	Expandable Closed Path Eddy Covariance System w/EC155 & Pump Module
CPEC306-6NA	Expandable Closed Path Eddy Covariance System w/EC155 & Pump Module, with CR6
CPEC310	Expandable Closed Path Eddy Covariance System w/EC155,Pump Module, & 3-Valve Module
CPEC310-6NA	Expandable Closed Path Eddy Covariance System w/EC155,Pump Module, & 3-Valve Module, with CR6

The object of the declaration described above is in conformity with the relevant Union harmonization legislation:

2014/30/EU The Electromagnetic Compatibility Directive (EMC)

2011/65/EU The Restriction of Hazardous Substances Directive (RoHS2)

Standards referenced during compliance assessment:

EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use – EMC requirements – for use in industrial locations.

EN 50581:2012 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances.

Signed for and on behalf of:

Place: Campbell Scientific Inc.

Date: February 15, 2018

Karsten Sorensen
Quality Assurance Manager
Campbell Scientific Inc.