



Declaration of Conformity
(in accordance with UK government legislation)

Product: ClariVUE10

Manufacturer:

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F
www.campbellsci.com

This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):

Model	Description
ClariVUE10	Turbidity Sensor W/ISO7027 Optics

The object of the declaration described above is in conformity with the relevant UK Statutory Instruments and their amendments:

- 2016 No. 1091** The Electromagnetic Compatibility Regulations 2016
- 2016 No. 3032** The Restriction of the Use of Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

Standards referenced during compliance assessment:

- BS EN 61326-2-3:2013** Electrical equipment for measurement, control and laboratory use – EMC requirements; Criteria for transducers with integrated or remote signal conditioning.
- BS EN 50581:2012** Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances.

Signed for and on behalf of:

Place: Campbell Scientific Inc.

Date: Jan 19, 2021

Name: Gary Travis Roberts

Title/Position: Manager, Product Test and Compliance

Tel: (435) 227-9623

E-mail: gtroberts@campbellsci.com