



**Declaration of Conformity**  
(in accordance with UK government legislation)

**Product: NL121**

**Manufacturer:**

Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | 435.227.9000 | 435.227.9001 F  
[www.campbellsci.com](http://www.campbellsci.com)

**This declaration of conformity is issued under the sole responsibility of the manufacturer for the following product(s):**

Model	Description
NL121	Ethernet Interface

**The object of the declaration described above is in conformity with the relevant UK Statutory Instruments and their amendments:**

- 2016 No. 1101** The Electrical Equipment Safety Regulations 2016
- 2016 No. 3032** The Restriction of the Use of Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

**Standards referenced during compliance assessment:**

- BS EN 61326-1:2013** Electrical equipment for measurement, control and laboratory use – EMC requirements - for use in industrial locations.
- BS EN 50581:2012** Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances.

**Signed for and on behalf of:**

**Place:** Campbell Scientific Inc.

**Date:** May 17, 2021

Name: Gary Travis Roberts

Title/Position: Manager, Product Test and Compliance

Tel: (435) 227-9623

E-mail: [gtroberts@campbellsci.com](mailto:gtroberts@campbellsci.com)